## Food and Drug Administration, HHS

parts. This generic type of device does not include devices used to generate the ultrasonic frequency electrical signals for application.

(b) Classification. Class II (performance standards).

## §884.2980 Telethermographic system.

- (a) Telethermographic system intended for adjunctive diagnostic screening for detection of breast cancer or other uses—(1) Identification. A telethermographic system for adjunctive diagnostic screening for detection of breast cancer or other uses is an electrically powered device with a detector that is intended to measure, without touching the patient's skin, the self-emanating infrared radiation that reveals the temperature variations of the surface of the body. This generic type of device may include signal analysis and display equipment, patient and equipment supports, component parts, and acces-
- (2) Classification. Class I (general controls)
- (b) Telethermographic system intended for use alone in diagnostic screening for detection of breast cancer or other uses-(1) Identification. A telethermographic system for use as the sole diagnostic screening tool for detection of breast cancer or other uses is an electrically powered device with a detector that is intended to measure, without touching the patient's skin, the self-emanating infrared radiation that reveals the temperature variations of the surface of the body. This generic type of device may include signal analysis and display equipment, patient and equipment supports, component parts, and accessories.
  - (2) Classification. Class III.
- (3) Date PMA or notice of completion of a PDP is required. As of the enactment date of the amendments, May 28, 1976, an approval under section 515 of the act is required before the device described in paragraph (b)(1) may be commercially distributed. See §884.3.
- [53 FR 1566, Jan. 20, 1988, as amended at 55 FR 48440, Nov. 20, 1990; 66 FR 46953, Sept. 10, 2001]

## §884.2982 Liquid crystal thermographic system.

- (a) A nonelectrically powered or an ACpowered liquid crystal thermographic system intended for adjunctive use in diagnostic screening for detection of breast cancer or other uses—(1) Identification. A nonelectrically powered or an AC-powered liquid crystal thermographic system intended for use as an adjunct to physical palpation or mammography in diagnostic screening for detection of breast cancer or other uses is a nonelectrically powered or an AC-powered device applied to the skin that displays the color patterns of heat sensitive cholesteric liquid crystals that respond to temperature variations of the surface of the body. This generic type of device may include patient and equipment supports, a means to ensure thermal contact between the patient's skin and the liquid crystals, component parts, and accessories.
- (2) Classification. Class I (general controls).
- (b) A nonelectrically powered or an ACpowered liquid crystal thermographic system intended for use alone in diagnostic screening for detection of breast cancer or other uses-(1) Identification. A nonelectrically powered or an AC-powered liquid crystal thermographic system intended for use as the sole diagnostic screening tool for detection of breast cancer or other uses is a nonelectrically powered or an AC-powered device applied to the skin that displays the color patterns of heat sensitive cholesteric liquid crystals that respond to temperature variations of the surface of the body. This generic type of device may include image display and recording equipment, patient and equipment supports, a means to ensure thermal contact between the patient's skin and the liquid crystals, component parts, and accessories.
  - (2) Classification. Class III.
- (3) Date PMA or notice of completion of a PDP is required. As of the enactment date of the amendments, May 28, 1976, an approval under section 515 of the act is required before the device described in paragraph (b)(1) may be commercially distributed. See §884.3.

[53 FR 1566, Jan. 20, 1988, as amended at 55 FR 48441, Nov. 20, 1990; 66 FR 46953, Sept. 10, 2001]